



CERTIFICATE OF CONFORMITY

Issuer's Name / Producer : GLASSCO LABORATORY EQUIPMENTS PVT. LTD.

Issuer's Address / Producer : Manglai, P.O. Khudda Kalan, Ambala Cantt - 133004 Haryana, India

Object of Declaration : Funnel

Material : Borosilicate Glass

Purpose of Use : it used to guard against spillage when pouring chemicals from one vessel to another, and they can also be fitted with a filter to separate solid form liquid.

Specification of funnel specified in Table-1, DIN ISO 4798

Catalogue No.	Product Description	Size (mm)	Pack Qty
238.202.01*	Funnel	25	10
238.202.02	Funnel	35	10
238.202.03*	Funnel	38	10
238.202.04*	Funnel	40	10
238.202.05*	Funnel	50	10
238.202.06	Funnel	55	10
238.202.07*	Funnel	65	10
238.202.08*	Funnel	70	10
238.202.09	Funnel	75	10
238.202.10*	Funnel	80	10
238.202.11	Funnel	100	10
238.202.12	Funnel	125	10
238.202.13	Funnel	150	10



*Non Din

The object of declaration described above is in conformity with the requirements of the following documents:

- DIN ISO 3585 Borosilicate Glass 3.3
- DIN ISO 4798 Laboratory Glassware - Funnel

Material: The above are made from DIN ISO 3585:1998, heat resistant Boro 3.3 Glass that meet ASTM E-438, TYPE-1, CLASS- A Standard & USP TYPE-1 Glass.

Compliance: These funnels comply in every respect characteristics, measures and accuracy are in full conformity With the provisions of the standard DIN ISO 4798:1997.

Additional Information:

Glassco has installed the quality control system as per ISO 9001 and products delivered to the market are in full conformity with the technical Specification as per standard and with all fundamental requirements to such products.

Certificate No. IPK 0039.06 issued by intertek certification limited, A UKAS Management System Accredited Body



This document has been produced electronically & is valid without signature